(b) *Classification*. Class II (performance standards).

# §880.5150 Nonpowered flotation therapy mattress.

- (a) *Identification*. A nonpowered flotation therapy mattress is a mattress intended for medical purposes which contains air, fluid, or other materials that have the functionally equivalent effect of supporting a patient and avoiding excess pressure on local body areas. The device is intended to treat or prevent decubitus ulcers (bed sores).
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807. The device also is exempt from the good manufacturing practice regulation in part 820, with the exception of §820.180, with respect to general requirements concerning records, and §820.198, with respect to complaint files.

#### §880.5160 Therapeutic medical binder.

- (a) *Identification*. A therapeutic medical binder is a device, usually made of cloth, that is intended for medical purposes and that can be secured by ties so that it supports the underlying part of the body or holds a dressing in place. This generic type of device includes the abdominal binder, breast binder, and perineal binder.
- (b) Classification. Class I (general controls). The device is exempt from the premarket notifiction procedures in subpart E of part 807. If the device is not labeled or otherwise represented as sterile, it also is exempt from the good manufacturing practice regulation in part 820, with the exception of §820.180, with respect to general requirements concerning records, and §820.198, with respect to complaint files.

#### §880.5180 Burn sheet.

- (a) *Identification*. A burn sheet is a device made of a porous material that is wrapped aroung a burn victim to retain body heat, to absorb wound exudate, and to serve as a barrier against contaminants.
- (b) Classification. Class I. The device is exempt from the premarket notifica-

tion procedures in subpart E of part 807 of this chapter.

[45 FR 69682-69737, Oct. 21, 1980, as amended at 59 FR 63011, Dec. 7, 1994]

#### §880.5200 Intravascular catheter.

- (a) *Identification*. An intravascular catheter is a device that consists of a slender tube and any necessary connecting fittings and that is inserted into the patient's vascular system for short term use (less than 30 days) to sample blood, monitor blood pressure, or administer fluids intravenously. The device may be constructed of metal, rubber, plastic, or a combination of these materials.
- (b) *Classification*. Class II (performance standards).

### §880.5210 Intravascular catheter securement device.

- (a) *Identification*. An intravascular catheter securement device is a device with an adhesive backing that is placed over a needle or catheter and is used to keep the hub of the needle or the catheter flat and securely anchored to the skin.
- (b) *Classification.* Class I. The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter.

[45 FR 69682-69737, Oct. 21, 1980, as amended at 59 FR 63011, Dec. 7, 1994]

## § 880.5240 Medical adhesive tape and adhesive bandage.

- (a) *Identification*. A medical adhesive tape or adhesive bandage is a device intended for medical purposes that consists of a strip of fabric material or plastic, coated on one side with an adhesive, and may include a pad of surgical dressing without a disinfectant. The device is used to cover and protect wounds, to hold together the skin edges of a wound, to support an injured part of the body, or to secure objects to the skin.
- (b) *Classification*. Class I. The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter.
- $[45\ FR\ 69682\text{-}69737,\ Oct.\ 21,\ 1980,\ as\ amended$  at  $59\ FR\ 63011,\ Dec.\ 7,\ 1994]$